

Certificate of Compliance

Certificate No. : BSTXD20020147810001Y-1EC-1

Applicant : Yaoyue Technology (Shenzhen) Co., Ltd

103, No. 14, Jinmao Road, Guancheng community, Guanhu street,

Longhua District, Shenzhen, Guangdong China

Manufacturer : Yaoyue Technology (Shenzhen) Co., Ltd

103, No. 14, Jinmao Road, Guancheng community, Guanhu street,

Longhua District, Shenzhen, Guangdong China

Product Name: : FINGERTIP PULSE OXIMETER

Main Test Model : AB-80

Additional Model : AB-88, AB-82, AB-85, AB-86

Test Standard

EN61326-1:2013

As shown in the

Test Report No.

BSTXD20020147810001Y-1ER-1

The EUT described above has been tested by us with the listed standards and found in compliance with the council EMC directive 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC Directive.

The certificate applies to the tested sample above mentioned only and shall not imply an assessment of the whole production.



CE



Apr. 09, 2020

Shenzhen BST Technology Co., Ltd.

Add: Building No.23-24, Zhiheng Industrial Park, Guankouer Road, Nantou, Nanshan District, Shenzhen, Guangdong, China Certificate Search: http://www.bst-lab.com. Tel:400-882-9628, 8009990305, E-mail: christina@bst-lab.com



Certificate of Compliance

Certificate Number: BST200414793701CC

Applicant : YAOYUE TECHNOLOGY (SHENZHEN) CO., LTD

103, No. 14, Jinmao Road, Guancheng Community, Guanhu Street,

Longhua District, Shenzhen, Guangdong China

Manufacturer : YAOYUE TECHNOLOGY (SHENZHEN) CO., LTD

103, No. 14, Jinmao Road, Guancheng Community, Guanhu Street,

Longhua District, Shenzhen, Guangdong China

Product Name : FINGERTIP PULSE OXIMETER

Main Test Model : AB-80

Additional Model : AB-88, AB-82, AB-85, AB-86

Test Standard : IEC 62321-4:2013+AMD1:2017, IEC 62321-5:2013,

IEC 62321-6:2015, IEC 62321-7-1:2015, IEC 62321-7-2:2017, IEC 62321-8:2017

As shown in the Test Report No.

: BST200414793701CR

The EUT described above has been tested by us and found in compliance with the council RoHS 2 Directive 2011/65/EU Annex II (EU) 2015/863 as last amended by Directive (EU) 2017/2102. This certificate is only valid in conjunction with the test report.



RoHS



Tony Qian Approved Signatory Apr.22, 2020

Dongguan BST Testing Co., Ltd

Add: A1201-1204 Xinsanqi of Dongbao Road, Dongcheng District, Dongguan, Guangdong, China Certificate Search: http://www.bst-lab.com, Tel: 400-8829628, 800-9990305, E-mail: christina@bst-lab.com



Supplier's Declaration of Conformity

This device complies with part 15 of the FCC Rules. Operation is subject to the condition that this device does not cause harmful interference

Certification Summary

Yaoyue Technology (Shenzhen) Co., Ltd. Applicant

03, No. 14, Jinmao Road, Guancheng community, Guanhu Address

street, Longhua District, Shenzhen, Guangdong China

Yaoyue Technology (Shenzhen) Co., Ltd Manufacturer

03, No. 14, Jinmao Road, Guancheng community, Guanhu Address

street, Longhua District, Shenzhen, Guangdong China

FCC Part 15 Class B Test standard

FINGERTIP PULSE OXIMETER Product Type

Main Test Model AB-80

AB-88, AB-82, AB-85, AB-86 Additional Model

Trade Name N/A

Certificate Number BSTXD20020147810002-1EC Report number BSTXD20020147810002-1ER

2020-04-09 Issuance Date

Shenzhen BST Technology Co., Ltd. Tested by

Signature of Officer

Christina Deng -Certification Manager

We the responsible party:

Yaoyue Technology (Shenzhen) Co., Ltd

declare that the product ADULT SUPPLIES





was tested to conform to the applicable FCC Rules and regulations. The method of testing was in accordance to the most accurate measurement standards possible, and that all necessary steps have been enforced to assure that all production units of the same equipment will continue to comply with the Federal Communications Commission's requirements.

Signature	Date	
Name	Title	

Shenzhen BST Technology Co.,Ltd.

Add: Building No.23-24, Zhiheng Industrial Park, Guarikouer Road, Nantou Nanshan District, Shenzhen, Guangdong, China Certificate Search: http://www.bst-lab.com. Tel:400-882-9828, 8009990305, E-mail:christina@bst-lab.com